SHREE ANALYTICAL TESTING & RESEARCH LABORATORY



Govt. of India, Ministry of Agricultural marketing & Inspection Dept approved "AGMARK" Food testing Lab.

Govt. of Maharashtra approved Soil & Water testing Laboratory (Nashik Region)

ISO 9001: 2008 Certified Company

Test Report

			. cot itopott		
1. 2.	Test Report No Issue Date	:-	Rep. No:- SATRL/ 22-23/03/181 12-03-2023		
3.	Name & Address of the Customer	:-	Shri Chamundamata Homoepathic Medi College & Hospital, jalgaon, Maharashtra		
4.	Reference	:-	As per Sample Testing Letter from Party		
5. 6. 7. 8.	Date of Sample Receipt Lab Id Sample Submitted By Page No.	1 1 1 1	05-03-2023 SATRL/ 22-23/03/174 Customer 1 of 1		

Sample Details

Client Ref. Letter No/Dt: NA

Drinking Water (Homeopathy College)

Sample Quantity: **Analysis Start Dt:** 1 LIT 05-03-2023

Sample ID /Name : Batch No /Supply:

Not Applicable

Analysis End Dt:

08-03-2023

Chemical Testing: -

Test Results

SR.NO.	TEST PARAMETERS	RESULT	IS 10500 STANDARD	TEST METHOD	
1	pH Value	7.13 51.30 μS	6.50 - 8.50 Not Specified	IS 3025 (P-11):2022	
2	Electrical Conductivity			IS 3025 (P-14)-1984	
3	Total Dissolved Solids	33.34	< 500 mg/l	IS 3025 (P-16):2023	
4	Chloride, as Cl	18.43	< 250 mg/l	IS 3025 (P-32):1983 RA 2019	
5	Total Alkalinity, as CaCO ₃	6.00	< 250 mg/l	IS 3025 (P-23):1986 RA 2019	
6	Calcium, as Ca	4.80	< 75 mg/l	IS 3025 (P-40):1991 RA 2019	
7	Magnesium, as Mg	0.58	< 30 mg/l	IS 3025 (P-46):1994 RA 2019	
8	Total Hardness, as CaCO ₃	12.00	< 200 mg/l	IS 3025 (P-21):1984 RA 2009	
9	Fluoride, as F	BLQ	< 1.0 mg/l	IS 3025 (P-60):2008	
10	Nitrate, as NO₃	BLQ	< 45 mg/l	APHA 4500 B 23 rd Ed	
11	Iron, as Fe	BLQ	< 0.30 mg/l	IS 3025 (P-2):2019	
12	Sulphate, as SO ₄	5.00	< 200 mg/l	IS 3025 (P-24/Sec 1):2022	
	M	licrobiologica	l Testing	MAN VARIABLE W	
SR.NO.	TEST PARAMETERS	RESULT	IS 10500 STANDARD	TEST METHOD	
1	Total coliforms	Absent	Absent in 100 ml Sample	IS 15185:2016 RA 202	
			TOTAL CONTRACTOR OF THE PARTY O	The same of the sa	

Absent ml Sample *Remark: - Water can be used for drinking purpose as it is within limits of BIS. (Bureau of Indian Standards). (BLQ-Below Limit of Quantification)

Absent in 100

Tested by

E.coli

IS 15185:2016 RA 2021

Authorized by

Authorized by

Technical Manager

Te Chemist Conditions of reporting Vikas October any legal action, unless prior permission only), unless specific instructions to the o

getable Market ,Ajintha Road, Jalgoan - 425 003 (MS.)